

# **BS-480 Chemistry Analyzer**



# **BS-480 Chemistry Analyzer**

# **Robust hardware**

## **Enhanced Liquid System**

- Precise sample and reagent aspiration
- Built in air bubble elimination prior to washing
- High pressure interior probe wash
- Carry-over < 0.05%



## **Economy Usage**

- Light-Spot Flatting Technology facilitates lower reaction volume
- 24 hour non-stop refrigeration at 2-10°C
- Minimum sample volume: 1.5 μl
- Reagent volume: 10 ~ 350 μl
- States Cuvette Volume:120~360µl





## Easy to perform maintenance

- Front loaded reagent compartment for easier access Debbie
- Easy access for routine maintenance and troubleshooting
- Built in Step-by-Step maintenance guide

# Intelligent Probe System functions and smart protection

- Supports methods with up to 4 reagents
- Vertical and horizontal collision protection
- Automatic System Recovery
- Liquid level detection, clot detection



# **Tailor made**



# for your lab

# **Advanced software**

	Sam		Quality Con	est 1	(Marco)				
) 	10 1		est 1	Pes 1		Sample Type	Deam 🗸		_
	Bar Code			0	heren		~	Patient ID	
Resalt		Ca.	- 49	. P.	Glant	0440	T-BAV	0.847	DH
(i) Raspert	7460	ADA	M	TBA	41	AST	ALP	104	Let.
	TP	-	ALB	UP	+ 1111	Apeld	40.8	10	LDL
e c	HOLIC	TC	CREAS	CREAJ	u.	Cysc	UNEA	C02	CK.4
-	CK.	LOH	++60H	P.N	HECOP	ASO		•	>
Alerr									
									_
	Demog F1	Cydoa F2	- Date	Ch		Jat	Prev	Net FZ	dave Fit

	54					•		
m	Levey-Jo	wings 🗍	Tuis-FM	Res	dis 1	Survey	Swip	
Program	M	M1	~					
_	Chem	ALB.	Show	Deleted	Fini Result			
	OC Date	2012- 1- 1	2012-11-00		Cone:		Cure Mean:	
Realt	Central 1	(0.0.0)		-	UNK		Sum SD:	
100	Compol 2	OMP2H		~	Fran Date:		Sun CV%:	
Teacers.	Compol 3	_		~	Comment		Gare N.	
	11111	***				100		
Alam	時時							

	Currer			Hatay	State	ics					
hogan	Ry Sample		7.01	<ul> <li>Ry Chevelary</li> <li>A Sample D Bar Code</li> </ul>	Pestion	Datus	Congletion	Read	Freq	RE Lot No.	
	AST	^	•	8001	141	Complete	11.00	12.0		1019	
Resalt	CRP				14	Complete	10.44	8.01	EDT.	1019	
			A		14	Complete	11.05	84		9019	
			×	1	10	Complete	10.44	85.0		9019	
Loogert			*	3	10	Complete	10.45	2.8		9019	
				4	14	Complete	10.45	8.6		9019	
			A	8	1.6	Complete	10.45	4.6		9019	
				•	14	Complete	10.45	11.4		9019	
0			κ.	7	17	Complete	10.45	15.0	REC	1019	
100					1.0	Complete	16.45	-0.2		9019	
_			×		1.0	Complete	10.45	3.6	4	9019	
				10	1-50	Complete	10.45	95.3		9019	
Alexa				91	1-11	Complete	10.45	43		9019	~
		V	<							>	

Program	Chersistry Reflec Relation	GRAH			Enable R Conditions	elles Function			
Feent	-	ione	Falded Chemistry		•		OK	*	
-				^	0 .		ad	•	
(B)					Related Chemi	aby			
	-			-	C.	Mg	•	0440	7.80
~				-1	D-BHV	0-68-0	T-68-D	ADA	PA
1					TEA	ALT	AST	ALP	¥6.
A					Local	тр	сня	ALE	UP
Alam				_	o-AMI	AppA1	Apolt	<	>
0	<			>	Culate	Car	-	CK	D.

Curre Workle	ed	History Results Obtivit: Data	Statistics		
Cheminky	AGT	Total Tests	12	Mean	82
50	5.479	Meximum	16.3	Meimun	4.2
Cont					
s : _					
					_
					-
				1 min 1	
• • <u>.</u>	14	0 0		41 87	62 54
( Seat)	1 11	14	26	20	26
Search	1 1			I I	Post

## **User-friendly Interface**

- Real-time analytical and carousel status monitoring
- Bi-directional LIS interface transmission

## **Real-time QC Status Monitoring**

- Levy-Jennings chart and Twin-Plot chart
- QC Out-of-Range real-time alarm
- Customizable periodic QC reminder

## **Traceable Test Results**

- Historic data recall from reagent/ calibrator/ control archive
- Intuitive software, intuitive software design, easy historical results

## **Reflex Function**

- User-defined customizable reflexive assays
- EMultiple reflexive criteria may apply to a single assay
- A single reflexive criteria may apply up to 20 assays

## **Test Summary**

- Test summary reports & sample rerun available on all assays and Quality Assurance
- Facilitate computation of total test costs
- Error Log Export function -facilitate error report to engineers
- Results Archive can be transferred to engineers for evaluation

# Accurate, Reliable Results

To ensure accuracy, reliability and correlation of diagnostic data, Mindray utilizes the International Standard in result reporting. To assure ease of report retrieving, Mindray establishes the Mindray Clinical Chemistry Measurement System for result traceability.


## Standard reference system

- Adopt JCTLM reference system
- IFCC primary method for enzyme, ID/MS method for substrate
- NIST, IRMM reference materials

JCTLM, Joint Committee On Traceability In Laboratory Medicine NIST, National Institute of Standards and Technology, USA IRMM, Institute for Reference Materials and Measurements, EU IFCC, International Federation of Clinical Chemistry and Laboratory Medicine


## **Complete traceability process**

 Complete calibration hierarchy and traceability chain based on ISO standard (EN/ISO17511) from reference system to routine measurement system



# Proficiency testing for reference measurement

• Participate RELA (External quality control for reference laboratory) to verify the accuracy of the value assignment procedure.









## International standardization certification

International Standardization certificates of Cholesterol and HbA1c from CRMLN and NGSP.

More information refers to website (http://www.cdc.gov).

## CRMLN (Cholesterol Reference Method of Laboratory Network) NGSP(National glycosylated hemoglobin standardization program)



## Matched calibrators and controls

- Dedicated calibrators with traceability and specific target value
- Convenient design of multi items of calibrators and controls combined into one vial
- Long shelf life of lyophilized powder

## Dedicated, high-quality reagents

• Diagnostic function test panels

Test panels such as : Hepatic panel, renal panel, pancreatic panel, lipid panel, cardiac panel, diabetic panel, rheumatic factor panel

Reliable analysis performance

EP series standard (CLSI)-evaluate and optimize reagent system for reliable performance in precision, linearity, stability, specificity and anti-interference capability

ISO standard manufacturing

uncentration:

Mindray follows straightly the ISO certified manufacturing process to ensure every lot of reagent in production is of supreme quality

ight)

# **Reagent Menu**

## **Hepatic Panel**

Alanine Aminotransferase (ALT) Aspartate Aminotransferase (AST) Alkaline Phosphatase (ALP) γ-GlutamylTransferase (γ-GT) Direct Bilirubin (D-Bil) DSA Method Direct Bilirubin (D-Bil)VOX Method Total Bilirubin (T-Bil) DSA Method Total Bilirubin (T-Bil)VOX Method Total Bilirubin (ALB) Total Bile Acids (TBA) Prealbumin (PA) Cholinesterase (CHE) a-L-fucosidase (AFU) 5'-nucleotidase (5'-NT)

## **Renal Panel**

Urea (UREA) Creatinine (CREA) Modified Jaffé Method Creatinine (CREA)Sarcosine Oxidase Method Uric Acid (UA) Carbon dioxide (CO2) Microalbumin β2-Microglobulin (β2-MG) Cystatin C (CysC) Retinol binding protein( RBP)

### Cardiac panel

Creatine Kinase (CK) Creatine Kinase-MB (CK-MB) Lactate Dehydrogenase (LDH) α-Hydroxybutyrate Dehydrogenase(α-HBDH) High sensitive C-reaction protein(HS-CRP) Homocysteine (HCY) Myoglobin(MYO) D-Dimer(D-Dimer)

## Inorganic & Anemia

Iron (Fe) Ferritin (FER) Transferrin (TRF) Calcium (Ca) Magnesium (Mg) Phosphate Inorganic (P) Unsaturated iron binding capacity (UIBC) Glucose-6-phosphate dehydrogenase (G6PD)

## Lipid Panel

Total Cholesterol (TC) Triglycerides (TG) HDL-Cholesterol (HDL-C) LDL-Cholesterol (LDL-C) Apolipoprotein A1 (ApoA1) Apolipoprotein B (ApoB) Lipoprotein(a) [Lp(a)]

### Immune Panel

Immunoglobulin A (IgA) Immunoglobulin G (IgG) Immunoglobulin M (IgM) Immunoglobulin E (IgE) Complement C3 (C3) Complement C4 (C4)

### **Diabetes Panel**

Glucose (Glu) GOD-POD Method Glucose (Glu) HK Method Hemoglobin A1c (HbA1c) Fructosamine (FUN) β-Hydroxybutyrate(β-HB)

### **Rheumatism Panel**

C-reactive protein (CRP) Rheumatoid Factor (RF) Antibodies Against Streptolysin O (ASO)

#### **Pancreatitis Panel**

α-Amylase (α-AMY) Lipase (LIP)

### Lung Panel

Adenosine Deaminase (ADA) Angiotensin Converting Enzyme(ACE)



# Mindray solution for clinical chemistry



Mindray can now provide 60 parameters of dedicated reagents (more than 8 others are coming), covering hepatic, renal, cardiac, lipids, diabetes, pancreatitis, inorganic ions and immunalassays, etc.,together with original calibrators with metrological traceability as well as controls for BS-480 chemistry analyzer.



# **BS-480 Chemistry Analyzer**

## **Technical Specifications**

#### System function

Fully automated,	discrete, random access,
STAT, urine and h	nomogeneous immunoassays;STAT sample priority
Throughput:	400 photometric tests/hour, up to 240 tests/hour for ISE
Measuring princi	ples: Absorbance Photometry, Turbidimetry
Methodology:	End-point, Fixed-time, Kinetic, ISE (Optional)
	Single/Dual/Triple/Quadruple reagent chemistries,
	Monochromatic/Bichromatic
Programming:	User defined profiles and calculation

#### Sample Handling

Sample tray:	90 positions for primary or secondary tubes and sample cups
Sample volume:	1.5~45 μl, step by 0.1μl
Sample probe:	Liquid level detection, clot detection and collision protection
Probe cleaning:	Interior and exterior automatic probe washing
	carry-over < 0.05%
	Automatic sample dilution, Pre-dilution and post-dilution
	Dilution with ratio up to 1: 150
Dilution vessel:	Quartz cuvette

#### Internal bar code reader

Sample/Reagent barcode reading including Codabar, ITF (Interleaved Two of Five), code128, code39, UPC/EAN, Code93; Bi-directional LIS Interface transmission

#### ISE Module (optional)

Optional selection of K<sup>+</sup>, Na<sup>+</sup>, Cl<sup>-</sup> Throughput: Up to 240 tests per hour

#### **Reagent Handling**

Reagent tray:	80 positions in refrigerated compartment (2~10°C)
Reagent volume:	10~350µl
Reagent probe:	Liquid level detection, collision protection and
	inventory check
Probe cleaning:	Interior and exterior automatic probe washing

#### Reaction System

Reaction rotor:Rotating tray, 90 cuvettes with automatic washingCuvette:Optical length 5mmReaction volume:120~360µlOperating temperature:37°CTemperature fluctuation:±0.1°CMixing system:2 independent mixers

#### **Optical System**

Halogen-tungsten lamp
Reversed optics, grating photometry
340nm, 380nm, 412nm, 450nm, 505nm, 546nm,
570nm, 605nm, 660nm, 700nm, 740nm, 800nm
e: 0~3.3Abs (10mm conversion)
0.0001Abs

#### **Control and Calibration**

Calibration mode	: Linear (one-point, two-point and multi-point),Logit
	-Log 4P, Logit-Log 5P, Spline, exponential, Polynomial,
	Parabola
Control rules:	Westgard multi-rule, Levy-Jennings, Cumulative sum check, twin plot

#### **Operation Unit**

Operation system: Windows 8 Interface: RS-232, Network Port, USB/ parallel port

#### Working Conditions

Power Supply:	200~240V, 50/60Hz, 1500VA
	or 110~130V, 60Hz, 1500VA
Temperature:	15~30°C
Humidity:	35~85%
Water consumption	on: ≤20L/hour, De-ionized water
Dimension:	1180mm x 710mm x 1150mm (W x D x H)
Weight:	300 Kg

Mindray Building, Keji 12th Road South, High-tech Industrial Park, Nanshan, Shenzhen 518057, P.R. China Tel: +86 755 8188 8998 Fax: +86 755 26582680 E-mail: intl-market@mindray.com www.mindray.com

Mindray is listed on the NYSE under the symbol"MR"

mindray | Institute with much are registered trademarks or trademarks owned by Shenzhen Mindray Bio-medical Electronics Co., LTD. © 2014 Shenzhen Mindray Bio-Medical Electronics Co., Ltd. All rights reserved. Specifications subject to changes without prior notice. P/N:ENG-BS-480-210285x8-20140605

